

宏明科技校正實驗室

Hong-Ming Technology Calibration Laboratory

校正報告書

Report of Calibration

報告編號(Report Number): CRXXXXXXXXXX

報告日期(Report Date): XXXXXXXXXX

申請者 Applicant	XXXXXXXXXXXXXXXXXXXXXXXXXX				
申請者地址 Address	XXXXXXXXXXXXXXXXXXXXXXXXXX				
儀器名稱 Equipment	XXXXXXXXXXXXXXXXXXXXXXXXXX				
製造商 Manufacturer	型號 Model No.	序號 Serial No.	收件日期 Received Date	校正日期 Calibration Date	校正地點 Calibration Location
HMT	MFS-630-TR	XXXXXXXXXXXXXX	XXXXXXXXXXXXXX	XXXXXXXXXXXXXX	實驗室

1. 上述委託儀器經本實驗室校正，結果如內文。

The calibration results of the above equipment calibrated by HMT Calibration Laboratory are indicated in the following pages.

2. 本校正報告含封面共 3 頁，分離視同無效。

This calibration report consists of 3 pages including the cover page. Single page is deemed invalid.

3. 未經本實驗室書面同意，本報告任何部分不得以任何形式或任何方式摘錄、複製、轉載或修改。

No part of this report may be excerpted, reproduced, duplicated, transferred, reproduced, and modified, in any form or by any means without the prior written permission of HMT Calibration Laboratory.

校正實驗室主管

Calibration Laboratory Manager

宏明科技校正實驗室

Hong-Ming Technology Calibration Laboratory

校正說明

Description of Calibration

報告編號(Report Number): CRXXXXXX

報告日期(Report Date): XXXXXXXX

實驗室環境條件 Environmental Conditions					
環境溫度(Temperature): (25 ± 5) °C					
相對溼度(Humidity): (60 ± 10) %					
校正用標準件 Calibration Standard					
標準件/型號 Certified Reference Material / Model No.	序號 Serial No.	追溯單位 Traceability	文件編號 File No.	校正日期 Calibration Date	有效日期 Expiration Date
Starna Neutral Density Filter Set/ RM-N1N35N	15315	UKAS Calibration Laboratory (0659)	5983	2015.10.02	2017.10.02

1. 校正程序依據 MFS-630 校正程序書(文件編號: HM-SOP-72-002)。

Calibration procedure is according to the MFS-630 calibration procedure (File No: HM-SOP-72-002).

2. 穿透率標準值 (%) = 標準件之穿透率數值

穿透率器示值 (%) = 送校件之穿透率量測值平均

穿透率器差值 (%) = 穿透率器示值-穿透率標準值

Transmittance standard values (%) = Transmittance values of Certified reference material

Transmittance measurement values (%) = Average of Transmittance measurement values

Transmittance deviation values (%) = Transmittance measurement values minus Transmittance standard values

3. 擴充不確定度計算依據 MFS-630 校正系統評估書(文件標號:HM-PM-72-07)。

擴充不確定度為組合標準量測不確定度與涵蓋因子($k = 2$ ，信賴水準約 95%)的乘積。

Expanded uncertainty is calculated from the MFS-630 calibration system evaluation (File No. HM-PM-72-07).

The reported expanded uncertainty is based on the combined standard measurement uncertainty multiplied by a coverage factor $k = 2$, providing a coverage probability of approximately 95%.

宏明科技校正實驗室

Hong-Ming Technology Calibration Laboratory

校正結果

Results of Calibration

報告編號(Report Number): CRXXXXXX

報告日期(Report Date): XXXXXXXX

波長(nm) Wavelength	穿透率標準值 (%) Transmittance Standard Values	穿透率器示值 (%) Transmittance Measurement Values	穿透率器差值 (%) Transmittance Deviation Values	擴充不確定度(%) Expanded Uncertainty
440	46.20	46.3	0.1	1.1
465	50.36	50.0	-0.4	1.1
546.1	50.06	49.9	-0.2	1.1
590	47.61	47.6	-0.1	1.1
635	47.11	47.1	-0.1	1.1
850	34.94	35.2	0.3	1.1
950	27.61	27.2	-0.4	1.1

(本頁以下空白 Null below)

校正者

Calibration performed by

報告簽署人

Approved Signatory